

194-204 = 10.146

204-64g = 0.043

206-208 = 2.004

Mount/sample No: ..... Date: ..... Page No: .....

Rejection over-ride	Sample/ Std ID	Time - printout	C1 2nd 254		Zr 20		U <sub>o</sub> 204Pb ppm	206 Kcps	206 Kcps	Y <sub>2</sub> Oucps	C1 2nd 206		C1 1st 207	Offsets OK?
			UO/U	196 Kcps	Age ±1σ (Ma)	206/238					207/206			
	MGI-1	14:13	11.6	0.6	2.6	6.8	1	55	0.884	0.063	wrong 208			
	XI-1.1	14:41	11.2	0.06	69.1	82	3	43	1.89	0.073	"			
	BSI-1	15:06	13.2	0.97	2.1	3	44	1.24	0.06	"				
	dv114el-1	15:40	7.2	0.2	2.1	2.2	2	27	1.36	0.13	✓			
	dv114gl-1	16:04	11.2	0.05	6.7	3.8	1	26	3.98	0.11	✓			
	dv114hl-1	16:32	12.3	0.07	9.8	6.1	8	353	3.97	0.112	✓			
	dv114hl-2	16:52	11.2	0.03	6.8	4.6	3	30	3.3	0.11	✓			
	dv114il-1	17:19	11.9	0.1	17.9	7.9	4	47	5.4	0.16	✓			
	dv114il-2	17:40	11.2	0.09	11.3	5.6	3	41	4.5	0.159	✓			
	MGI-2	18:01	12.3	0.4	3.7	9.9	1	51.5	0.91	0.059	✓			
	MGI-1-3	18:19	13.2	0.6	2.6	6.3	5	49	1.07	0.056	✓			
	BSI-2	18:38	14.5	1.4	1.0	2.0	2	43	1.38	0.068	✓			
	XI-1-2	18:57	12.4	0.05	7.5	79.3	1	45	2.36	0.073	✓			
	dv114cl-1	19:32	10.9	0.1	6.3	4.3	3	30.6	(3.2)	0.12	✓			
	dv114cl-2	19:56	12.2	0.1	4.8	2.7	3	25	4.3	0.12	✓			
	dv114cl-3	20:18	8.98	0.2	4.4	3.6	1	21	2.2	0.12	✓			
	dv114dl-1	20:51	11.3	0.06	4	1.96	6	10.5	4.8	0.13	✓			
	dv114dl-2	21:11	11.3	0.07	3	2.1	0	19.8	3.2	0.12	✓			
	dv114bl-1	21:42	10.3	0.2	5.8	5.1	6	25.8	2.3	0.108	✓			
	dv114bl-2	22:03	10.8	0.08	4.1	2.5	4	31	3.5	0.11	✓			
	MGI-4	22:22	12.3	0.4	3.4	8.2	0	53	1.00	0.059	✓			
	XI-1-3	22:41	7.33	0.04	84	85.5	6	44	2.6	0.073	✓			
	dv114al-1	23:09	10.7	0.2	11.1	8.2	4	28	2.88	0.115	✓			
	dv114al-3	23:33	7.5	0.3	3.8	4.1	1	22	1.4	0.13	✓			
	dv114el-2	00:01	11.8	0.12	5.6	3.6	2	31	3.7	0.12	✓			

fixed 208 trim mass value

Zr 0.007 high ratio



204-54pd = 0.043  
 206-208 = 2.004

Mount/sample No: .....

Date: .....

Rejection over-ride	Sample/Std ID	Time - printout	UO/U <del>20</del> 28	196 Kcps	206 cps kcps	Uo ppm ppb	204pb ppb	f206 %	Age ±1σ (Ma) 206/238 <del>25128</del>	207/206 <del>20721</del>	Offsets OK? 1st 207
	MGI-1-5	00:25	13.4	0.6	3.4	7.6	1	53	1.19	0.06	✓
	BSI-1-3	00:44	11.4	1.2	0.8	1.7	0	43	1.04	0.068	✓
	XI-1-4	01:03	12.8	0.05	73	73	2	39.8	2.5	0.072	✓
Sample Change HV off	MGI-1-6	01:41	12.1	0.6	3.1	7.7	1	56	0.98	0.053	inc. 193 by 0.002 ✓
	BSI-1-4	02:05	12.6	1.5	0.8	1.8	4	45	1.2	0.077	✓
	XI-1-5	02:25	12.9	0.05	80.5	84	1	42	2.48	0.072	✓
Zr high ←	dv113hl-1	02:57	9.6	0.5	2.9	4.9	2	14.9	1.12	0.127	✓
	dv113dl-1	03:21	9.04	0.3	4.9	5.4	10	25	1.66	0.11	✓
	dv113fh-1	03:49	5.95	0.07	1.5	2.2	1	8.0	0.82	0.138	✓
	dv113gh-1	04:27	10.2	0.13	6.1	4.4	1	22	2.86	0.119	✓
labelled it dv113fg and dv113fg	dv113gl-2	04:49	11.7	0.15	5.6	4.1	4	19.4	3.14	0.124	✓
	MGI-1-7	05:23	12.2	0.3	3.7	4.2	0	49	0.96	0.059	✓
	MGI-1-8	05:43	13.1	0.4	2.6	6.4	1	49	1.06	0.06	✓
	XI-1-6	06:03	12.6	0.05	69.7	66	4	39	2.65	0.072	✓
end session											